

**Certificate of Analysis****PRODUCT:** [REDACTED] Level 1 Modified Rabbit Plasma**CODE NO:** 31162L1**LOT NO:** 18837-B**DATE OF  
MANUFACTURE:** July 7, 2015**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.**EC 1069/2011:** AR-TEC-0007**EU  
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31162L1.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of contagious diseases.

**APPEARANCE:** Clear, amber fluid**ASSAYS:**

| Required Assays | Specification | Result |
|-----------------|---------------|--------|
| INR             | 1.1 – 1.2     | 1.2    |

| Meter     | Final Bulk | QC Testing<br>(sample 1) | QC Testing<br>(sample 2) | Average QC Testing |
|-----------|------------|--------------------------|--------------------------|--------------------|
| UP0083793 | 1.2        | 1.2                      | 1.2                      | 1.2                |
| UP0123048 | 1.2        | 1.2                      | 1.2                      | 1.2                |
| UP0092785 | 1.2        | 1.2                      | 1.2                      | 1.2                |
| UP0121390 | 1.2        | 1.2                      | 1.2                      | 1.2                |
| UP0122867 |            | .2                       | 1.2                      | 1.2                |

INR obtained using [REDACTED] meter.  
Test strips: code 08

**QUALITY ASSURANCE**

Quality Control Release Date: July 8, 2015

  
Sarah Graves  
Quality Systems Technician**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**