

Certificate of Analysis

PRODUCT: [REDACTED] Level 2 Modified Rabbit Plasma**CODE NO:** 31162L2**LOT NO:** 05538**DATE OF
MANUFACTURE:** February 24, 2016**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM;
-70 °C or below for 12 months from DOM.**EU 142/2011:** AR-TEC-0007**HEALTHY
ANIMAL
STATEMENT:** This product is made from raw materials that were collected from animals found free from obvious signs of infectious and contagious disease.**EU
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31162L2.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of contagious diseases.

APPEARANCE: Clear, amber fluid

Required Assays	Specification	Result
INR	3.1 – 3.3	3.2

Meter	Final Bulk	QC Testing (sample 1)	QC Testing (sample 2)	Average QC Testing
UP0078546	3.2	N/A	N/A	N/A
UP0010644	3.2	3.2	3.2	3.2
UP0076807	3.3	3.1	3.2	3.2
UP0413267	3.2	3.1	3.1	3.1
UP0755158		.1	3.2	3.2

INR obtained using [REDACTED] meter.
Test strips: code 08

QUALITY ASSURANCE

Quality Control Release Date: February 25, 2016

Sarah Graves
Quality Systems Technician**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**