

Certificate of Analysis

PRODUCT: [REDACTED] Level 2 Modified Rabbit Plasma**CODE NO:** 31162L2**LOT NO:** 21737**DATE OF
MANUFACTURE:** August 5, 2015**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM;
-70 °C or below for 12 months from DOM.**EU 142/2011:** AR-TEC-0007**HEALTHY
ANIMAL
STATEMENT:** This product is made from raw materials that were collected from animals found free from obvious signs of infectious and contagious disease.**EU
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31162L2.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of contagious diseases.

APPEARANCE: Clear, amber fluid

ASSAYS:	Required Assays	Specification	Result
	INR	31.-3.3	3.2

Meter	Final Bulk	QC Testing (sample 1)	QC Testing (sample 2)	Average QC Testing
UP0123048	3.3	3.2	3.1	3.2
UP0122867	3.3	3.2	3.1	3.2
UP0085453	3.2	3.2	3.1	3.2
UP0121390	3.2	3.2	3.1	3.2
UP0092785		.2	3.1	3.2

INR obtained using [REDACTED] meter.
Test strips: code 08

QUALITY ASSURANCE

Quality Control Release Date: August 6, 2015

Doug Stocker
Quality Systems Technician**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**