

rtificate of Analysis**PRODUCT:** [REDACTED] Level 2 Modified Rabbit Plasma**CODE NO:** 31162L2**LOT NO:** 21939**STORAGE:** -10 °C to -20 °C for 15 days from
DOM; -70 °C or below for 12
months from DOM.**DATE OF
MANUFACTURE:** August 7, 2017**ORIGIN:** U.S.A.**EU 142/2011:** AR-TEC-0007**HEALTHY
ANIMAL
STATEMENT:** This product is made from raw materials that were collected from animals found free
from obvious signs of infectious and contagious disease.All specimens that contain materials of animal origin were procured using the latest
animal testing guidelines, laws or regulations and materials were obtained from US
Department of Agriculture ("USDA") approved facilities.**EU
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez
Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using
Pel-Freez SMP 31162L2.Rabbit blood was collected and processed from New Zealand white rabbits
(mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were
found to be free of contagious diseases.**APPEARANCE:** Clear, amber fluid**ASSAYS:**

Required Assays	Specification	Result
INR	3.1 – 3.3	3.1

Meter	Final Bulk	QC Testing (sample 1)	QC Testing (sample 2)	Average QC Testing
UP0077741	3.2	3.0	3.1	3.1
UP0124142	3.2	3.0	3.1	3.1
UP0084011	3.1	3.0	3.1	3.1
UP0083793	3.2	3.1	3.1	3.1
UP0095119		.0	3.1	3.1

INR obtained using [REDACTED] meter.
Test strips: code X 10**QUALITY ASSURANCE**

Quality Control Release Date: August 11, 2017

Sarah Graves
Quality Systems Technician**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**