

alysis
PRODUCT: [REDACTED] Rabbit Plasma with EDTA, Sterile, Custom

CODE NO: 31164

LOT NO: 06437

PO#: [REDACTED]

DATE OF MANUFACTURE: March 5, 2015

Origin: U.S.A.

STORAGE: -20 °C

EU REQUIREMENTS: Product meets European Union requirements for production of technical blood products.

HEALTHY ANIMAL STATEMENT: This product is made from raw materials that were collected from animals and found free from infectious and contagious disease.

PROCEDURE: Rabbit whole blood is collected from healthy young rabbits (New Zealand White, 8 12 weeks of age, either sex, and fasted) at the time of slaughter combined with 25 mL EDTA then QS to 250 mL with blood. The plasma is separated from the blood cellular components by centrifugation and pooled. The specified amount of Sodium Chloride is added and allowed to mix until dissolution is complete. The pH is adjusted and the plasma is then passed through a 0.22-micron filter, bottled and stored at -20 °C.

ASSAYS:	Required Assays	Specification	Result
	Total Protein Assay:	4.0 – 6.5 g/dL	5.10 g/dL
	Hemoglobin Assay:	≤ 20 mg/dL	14.40 mg/dL
	Adjusted pH:	FIO	6.70 before sterile filtration
	Final pH:	FIO	6.88 @ 50% aqueous solution
	Bacterial Coagulase:	5 of 5 <i>S. aureus</i> strains +3 or greater at 6 - 24 hours. <i>S. epidermidis</i> & <i>E. aerogenes</i> negative at 6 - 24 hours.	5 of 5 <i>S. aureus</i> strains +3 or greater at 6 hours. <i>S. epidermidis</i> & <i>E. aerogenes</i> were negative at 6 hours.
	Optical Density:	0.250 – 0.500	0.319
	Sterility:	No growth at 14 days	Passed. No growth at 14 days.

QUALITY ASSURANCE

Quality Control Release Date: March 24, 2015



Sarah Graves

Quality Systems Technician

Meets or exceeds specifications ☒
Does not meet specifications ☐
SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES