

Certificate of Analysis**PRODUCT:** [REDACTED] Level 1 Modified Rabbit Plasma I.N.R. 1.0 – 1.2**CODE NO:** 31171**LOT NO:** 16047**DATE OF
MANUFACTURE:** June 9, 2025**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.**HEALTHY
ANIMAL
STATEMENT:** This product is made from raw materials that were collected from animals and found free from obvious signs infectious and contagious disease.**EU
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**EC 1069/2011:** AR-TEC-0007**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31171.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of obvious signs of contagious diseases.

APPEARANCE: Clear, amber fluid

ASSAYS:	Required Assays	Specification	Result
	INR	1.0 – 1.2	1.1

Meter	Final Bulk	QC Testing (sample 1)	QC Testing (sample 2)	Average QC Testing
UP01583599	1.1	1.2	1.1	1.2
UP01656560	1.1	1.1	1.1	1.1
UP01668059	1.1	1.2	1.1	1.2
UP1046206	1.1	1.1	1.1	1.1
UP01570790	1.1	1.1	1.1	1.1

INR obtained using [REDACTED] meter. Key code: S_486 XSPT
Test strips: code S_486 lot 560 993 80**QUALITY ASSURANCE**

Quality Control Release Date: June 13, 2025



Bartholomew Rothrauff

Quality Systems Technician

Meets or exceeds specifications ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**