

**State of Analysis****PRODUCT:** [REDACTED] Level 1 Modified Rabbit Plasma I.N.R. 1.0 – 1.2**CODE NO:** 31171**LOT NO:** 17845**DATE OF  
MANUFACTURE:** June 27, 2023**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.**HEALTHY  
ANIMAL  
STATEMENT:** This product is made from raw materials that were collected from animals and found free from obvious signs infectious and contagious disease.**EU  
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**EC 1069/2011:** AR-TEC-0007**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31171.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of obvious signs of contagious diseases.

**APPEARANCE:** Clear, amber fluid**ASSAYS:**

| Required Assays | Specification | Result |
|-----------------|---------------|--------|
| INR             | 1.0-1.2       | 1.2    |

| Meter      | Final Bulk | QC Testing (sample 1) | QC Testing (sample 2) | Average QC Testing |
|------------|------------|-----------------------|-----------------------|--------------------|
| UP01732892 | 1.2        | 1.2                   | 1.2                   | 1.2                |
| UP01656560 | 1.2        | 1.2                   | 1.2                   | 1.2                |
| UP01570790 | 1.2        | 1.2                   | 1.2                   | 1.2                |
| UP1229542  | 1.2        | 1.2                   | 1.2                   | 1.2                |
| UP01668059 | 1.2        | 1.2                   | 1.2                   | 1.2                |

INR obtained using [REDACTED] meter. Key code: S 486 XSPT  
Test strips: code S [REDACTED] 0 993 80**QUALITY ASSURANCE**

Quality Control Release Date: June 29, 2023



Doug Stocker

Quality Systems Technician

Meets or exceeds specifications ☒Does not meet specifications ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**