

**Certificate of Analysis****PRODUCT:** [REDACTED] Level 2 Modified Rabbit Plasma I.N.R. 3.2 - 3.4**CODE NO:** 31172**LOT NO:** 25043**DATE OF  
MANUFACTURE:** September 7, 2021**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.**HEALTHY  
ANIMAL  
STATEMENT:**

This product is made from raw materials that were collected from animals and found free from obvious signs of infectious and contagious disease.

**EU  
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**EC 1069/2011:** AR-TEC-0007**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31172.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of obvious signs of contagious diseases.

**APPEARANCE:** Clear, amber fluid**ASSAYS:**

| Required Assays | Specification | Result |
|-----------------|---------------|--------|
| INR             | 3.2-3.4       | 3.4    |

| Meter     | Final Bulk | QC Testing (sample 1) | QC Testing (sample 2) | Average QC Testing |
|-----------|------------|-----------------------|-----------------------|--------------------|
| UP0973274 | 3.4        | 3.3                   | 3.4                   | 3.4                |
| UP0086894 | 3.4        | 3.4                   | 3.4                   | 3.4                |
| UP0413267 | 3.3        | 3.4                   | 3.3                   | 3.4                |
| UP0870035 | 3.3        | 3.2                   | 3.4                   | 3.3                |
| UP0479147 | 3.4        | 3.4                   | 3.3                   | 3.4                |

INR obtained using [REDACTED] meter. Key code: S-385 XSPT  
Test strips: lot 393**QUALITY ASSURANCE**

Quality Control Release Date: September 10, 2021

Doug Stocker  
Quality Systems TechnicianMeets or exceeds specifications ☒Does not meet specifications ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**