

State of Analysis**PRODUCT:** [REDACTED] Level 3 Modified Rabbit Plasma I.N.R. 2.3 - 2.5**CODE NO:** 31173**LOT NO:** 29546**DATE OF
MANUFACTURE:** October 21, 2024**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.**HEALTHY
ANIMAL
STATEMENT:** This product is made from raw materials that were collected from animals and found free from obvious signs of infectious and contagious disease.**EU
REQUIREMENTS:** Product meets EU requirements for production of technical blood products. **EC 1069/2011:** AR-TEC-0007**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31173.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of obvious signs of contagious diseases.

APPEARANCE: Clear, amber fluid

ASSAYS:	Required Assays	Specification	Result
	INR	2.3 - 2.5	2.4

Meter	Final Bulk	QC Testing (sample 1)	QC Testing (sample 2)	Average QC Testing
UP01656560	2.5	2.4	2.4	2.4
UP1046206	2.5	2.4	2.4	2.4
UP1229542	2.4	2.4	2.4	2.4
UP01583599	2.4	2.4	2.3	2.4
UP01668059	2.4		2.4	2.4

INR obtained using [REDACTED] meter. Key code: S_486
Test strips: code S 80**QUALITY ASSURANCE**

Quality Control Release Date: October 24, 2024


Doug Stocker
Quality Systems Technician**Results pending** ☐**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**