

**rtificate of Analysis****PRODUCT:** [REDACTED] Level 2 Modified Rabbit Plasma I.N.R. 5.5-5.9**CODE NO:** 31184**LOT NO:** 05746**DATE OF  
MANUFACTURE:** February 26, 2024**ORIGIN:** U.S.A.**STORAGE:** -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.**HEALTHY  
ANIMAL  
STATEMENT:** This product is made from raw materials that were collected from animals and found free from obvious signs of infectious and contagious disease.**EU  
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**EC 1069/2011:** AR-TEC-0007**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31172.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 and were found to be free of obvious signs of contagious diseases.

**APPEARANCE:** Clear, amber fluid

| <b>ASSAYS:</b> | <b>Required Assays</b> | <b>Specification</b> | <b>Result</b> |
|----------------|------------------------|----------------------|---------------|
|                | INR                    | 5.5 – 5.9            | 5.5           |

| <b>Meter</b> | <b>Final Bulk</b> | <b>QC Testing<br/>(sample 1)</b> | <b>QC Testing<br/>(sample 2)</b> | <b>Average<br/>QC Testing</b> |
|--------------|-------------------|----------------------------------|----------------------------------|-------------------------------|
| UP1229542    | 5.6               | 5.5                              | 5.5                              | 5.5                           |
| UP01570790   | 5.5               | 5.7                              | 5.5                              | 5.6                           |
| UP01670440   | 5.7               | 5.6                              | 5.5                              | 5.6                           |
| UP01732892   | 5.6               | 5.6                              | 5.4                              | 5.5                           |
| UP01668059   | 5.7               | 5.4                              | 5.5                              | 5.5                           |

INR obtained using [REDACTED] meter. Key code: S\_486 XSPT  
Test strips: lot 560**QUALITY ASSURANCE**

Quality Control Release Date: February 28, 2024

Kendra Graham  
Quality Systems Technician / Clerk**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**