

Certificate of Analysis

PRODUCT: Rabbit Brain Thromboplastin Powder (**For PT Test**)**CODE NO:** 41103-0**LOT NO:** 15940**DATE OF
MANUFACTURE:** May 3, 2018**DATE OF
EXPIRATION:** May 3, 2023**STORAGE:** -10°C or below**ORIGIN:** U.S.A.**EU REQUIREMENTS:** Materials meet the requirements of the Regulation (EU) 142/2011 Chapter 8 Health Certificate.**EU 142/2011:** AR TEC 0007**HEALTHY
ANIMAL
STATEMENT:** This product is made from raw materials that were collected from animals and found free from obvious signs of infectious and contagious disease.**PROCEDURE:** Acetone-dried rabbit brain is composed solely of coarsely ground rabbit brain tissue, which has been dried using acetone. The rabbits (New Zealand White, 8-12 weeks of age) received both ante and post mortem inspection. The powder is buff to tan in color and is a coarse granular material that possesses a salty odor.

ASSAYS:	Required Assays	Specification	Result
	PT (Start4)	9.5 – 12.5 seconds	11.2 seconds
	Moisture (Halogen)	< 2.0 %	0.8%

QUALITY ASSURANCE

Quality Control Release Date: June 15, 2018



Sarah Graves

Quality Systems Technician

Amended by S.Graves on 22June2018: Description changed to Rabbit Brain Thromboplastin Powder (**For PT Test**)**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**