

Certificate of Analysis

PRODUCT: [REDACTED] Level 7 Modified Rabbit Plasma INR 7.7**CODE NO:** 60005**LOT NO:** 16341**DATE OF****MANUFACTURE:** June 12, 2019**STORAGE:** 15 days at -10°C to -20°C
or 12 months at -70°C or
below**ORIGIN:** U.S.A.**EU 142/2011:** AR-TEC-0007**EU****REQUIREMENTS:** Product meets EU requirements for production of technical blood products.**HEALTHY****ANIMAL STATEMENT:** This product is made from raw materials that were collected from animals found free from obvious signs of infectious and contagious disease.**PROCEDURE:** Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162).

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant P-202 .

ASSAYS:

Required Assays	Specification	Result
INR	7.7	7.9

Meter	Final Bulk	QC Testing (sample 1)	QC Testing (sample 2)	Average QC Testing
UP0087106	7.7	8.0	7.8	7.9
UP0852716	7.6	8.1	7.9	8.0
UP0870087	7.9	7.8	7.8	7.8
UP0497782	7.7	7.9	7.8	7.9
UP0908631	7.8	7.9	7.8	7.9

INR obtained using [REDACTED] meter. Key code: S-307
Test strips: lot 286**QUALITY ASSURANCE**

Quality Control Release Date: June 17, 2019

Doug Stocker
Quality Technician**Meets or exceeds specifications** ☒**Does not meet specifications** ☐**SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSES**