

Biologicals Division

219 N. Arkansas St.
205 North Arkansas
Rogers, Arkansas 72756

(479) 636-4361
(800) 643-3426

Fax 479-636-3562

Pel-Freez®

Analysis Certificate

PRODUCT: [REDACTED] Rabbit Plasma Level 3 Test Batch INR 2.2

CODE NO: 60010 **LOT NO:** 26735

STORAGE: -10°C to -20°C for 15 days from DOM; -70°C or below for 12 months from DOM.

**DATE OF
MANUFACTURE:** September 24, 2013

APPEARANCE: Clear, amber fluid

EC 1069/2009: EU # AR-TEC-0007

**EU
REQUIREMENTS:** Product meets EU requirements for production of technical blood products.

PROCEDURE: Modified rabbit plasma is produced from custom rabbit plasma (Pel-Freez Part # 91162) and adsorbed rabbit plasma (Pel-Freez Part # 81162), using Pel-Freez SMP 31162L3 to a fixed INR value as part of an ongoing robust assurance trial.

Rabbit blood was collected and processed from New Zealand white rabbits (mixed gender, 8-12 weeks, US grown) at USDA plant #P-202 and were found to be free of contagious diseases.

ASSAYS: INR obtained using [REDACTED] meter using test strips of code: 501, lot: 220 744 80. The QC Testin [REDACTED] page 2 of this certificate illustrates the ratios tested during production and final QC along with INR values.

QC Testing Data Table

Code: 60010 INR 2.2				
Lot#: 26735	Testing Stage:	In-process	Final QC	Final QC
DOM: 9/24/2013	Testing Sample Identification:	Post-bulk 1 mL test sample	Replicate #1 5 mL QC sample	Replicate #2 5 mL QC sample
Test strip lot: 220 744 80	Ratio (%) Custom Rabbit Plasma:Adsorbed Rabbit Plasma	28:72	28:72	28:72
Product Tested in Quintuplicate	Meter 1: UP0610760	2.2	2.2	2.2
	Meter 2: UP0734328	2.2	2.2	2.2
	Meter 3: UP0744403	2.2	2.2	2.2
	Meter 4: UP0570781	2.3	2.2	2.2
	Meter 5: UP0690781	2.2	2.2	2.2
Summary Statistics	Mean	2.2	2.2	2.2
	Std Deviation	0.045	0.000	0.000
	%CV	2.01%	0.00%	0.00%

Target INR – 2.2 INR

QUALITY ASSURANCE

Quality Control Release Date: October 2, 2013



Stephanie Whitmore
Quality Systems Technician

SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSE