

Pel-Freez®

Analysis Certificate

PRODUCT: Level 4 Modified Rabbit Plasma INR 6.6

CODE NO: 60015

LOT: 35235

STORAGE: -10 °C to -20 °C for 15 days from DOM; -70 °C or below for 12 months from DOM.

**DATE OF
MANUFACTURE:** December 18, 2013

**EU
REQUIREMENTS:** Product meets European Union requirements for production of technical blood products.

PROCEDURE: Rabbit blood was collected at USDA plant P-202 from New Zealand White rabbits (mixed gender, 8-12 weeks, US grown) that were found to be free of contagious diseases. Custom Rabbit Plasma was produced by mixing blood with custom anticoagulant solution according to Pel-Freez SMP 91162, chilled for 2 hours on wet ice and frozen at -70 °C. Adsorbed Rabbit Plasma was produced by mixing Custom Rabbit Plasma with barium sulfate, centrifuged, chilled and frozen at -70°C in the same manner.

Custom Rabbit Plasma and Adsorbed Rabbit Plasma were thawed, mixed with 9.972 g/L HEPES hemisodium salt, and blended together in proportions to yield an activated clotting time 6.6 INR.

After mixing thoroughly, the product was chilled for 2 hours on wet ice and frozen at -70°C.

The QC Testing Data Table on page 2 of this certificate illustrates the INR values of final QC testing.

EC 1069/2009:

EU 142/2011: EU # AR-TEC-0007

APPEARANCE: Clear, amber fluid

ASSAYS: INR obtained using [REDACTED] meter.
Test strips: code 5

Biologicals Division

219 North Arkansas St
205 North Arkansas
Rogers, Arkansas 72756

(479) 636-4361 Fax 479-636-3562
(800) 643-3426

Pel-Freez[®]

Analysis Certificate

QC Testing Data Table

Meter	1	2	Average
UP0744403	6.5	6.4	6.5
UP0734328	6.5	6.4	6.5
UP0570781	6.5	6.4	6.5
UP0610760	6.4	6.3	6.4
UP0751548	6.5	6.4	6.5

Target INR – **6.6** INR

QUALITY ASSURANCE

Quality Control Release Date: **January 9, 2014**



Sarah Graves
Quality Systems Technician

SUITABLE FOR FURTHER MANUFACTURE OR RESEARCH PURPOSE