

PRODUCT DATASHEET

FLUOROSHIELD™ (FLS) Mounting Medium**Catalog number: AR-6500-01**

AR-6500-02

AR-6500-03

AR-6500-05

AR-6500-01 AR-6500-02 AR-6500-03 AR-6500-05

Volume Volume Volume Volume

Component

FLUOROSHIELD™ (FLS) 20 mL 250 mL 1000 mL

100 mL

Mounting Medium Dropper bottle, Amber bottle Amber bottle

Amber bottle tip, and white cap

Description: Fluoroshield is an aqueous mounting medium for preserving fluorescence of tissue and cell smears. This unique formula prevents rapid photobleaching of FITC, Texas Red, AMCA, Cy2, Cy3, Cy5,

Alexa fluoro 488, Alexa fluoro 594, Green fluorescent protein (GFP), tetramethyl rhodamine, Redox, phycoerythrin (R-PE), phyocyanin (PC), and allophycocyanin (APC). The fluorescence is retained during prolonged storage at 4oC in the dark. This medium does not contain phenylenediamine, which destroys immunofluorescence of Cy dyes, R-PE, PC and APC.

Refractive Index: 1.34 -1.40.

Measured Refractive Index (RI) of the mounting media in liquid form was confirmed to be between

1.34-1.40. Refractive indexes change when the water solvent evaporates, and mounting media dries on slides. We do not have the means to measure the refractive indexes of dry mounting mediums; however, we expect the numbers to go higher when dried. RI of water is 1.333, RI of glass is 1.5.

Intended Use: Immunofluorescence, confocal microscopy.**Reagent: Ready to use mounting medium**

Storage: This product can be stored at room temp, however for long term storage 2-8oC is recommended.

Protect from light, DO NOT FREEZE.

May encounter problems with FROZEN BRAIN OR OTHER FROZEN TISSUES WITH LOTS OF

FAT**Procedure**

- 1. Bring mounting medium vial to room temperature.
- 2. Rinse slide to be mounted with DISTILLED OR DEIONIZED WATER, touch the edges of slide on a paper towel to remove excess water. Place slides on a flat surface away from light.
- 3. Turn the vial upside down and open the dropper to remove any air bubbles.
- 4. Apply 1-3 drops of mounting medium directly on top of the specimen.
- 5. Let stand at room temperature for about 3-5 minutes in the dark.

- 6. Apply cover slip carefully avoiding air bubbles.
- 7. The specimen is ready for visualization under a microscope.
- 8. One can seal the edges of cover slip with nail polish, any organic medium or our Organo mounting medium (cat. # AR-6504). If a coverslip is not sealed air bubbles will appear in few days.
- 9. Method for applying Coverslip: Put 1-3 drops of FLS on the specimen. After 3-4 minutes apply coverslip carefully avoiding air bubbles. Put Kimwipes on the top of coverslip, press gently to remove excess mounting medium. With 200 micropipette add organic mounting medium to seal the edges. Incubate at 37C for one hour in the dark to dry organic mounting medium.
- 10. For storage it is recommended that the slide be stored in the dark at 2-8oC.

Removal of Coverslip: Coverslip can be removed before sealing the edges. Soak slide in warm (37oC) distilled or deionized water for several minutes. Carefully and slowly move the coverslip. Soak in water for an additional few minutes to remove coverslip. Rinse slide several times with warm water to remove all mounting medium. The slide can be remounted again.

Limitation and warranty: Our warranty is limited to the actual price paid for the product. We are not liable for any property damage, personnel injury, time, effort or economic loss due to our product.

Safety: This product contains sodium azide as a preservative, appropriate care should be taken in handling. National Institute of Occupational Safety and Health has warning that sodium azide can react with lead, copper, brass or solder in the plumbing system and forms hydrazoic acid in acidic condition.

Discard with copious amount of water. Avoid skin and eye contact with all laboratory products. Use appropriate laboratory gear, lab coat, gloves and safety glasses. Do not ingest any laboratory products.

This product is not approved for administration in human or animals.

Country of Origin: USA

Fluoroshield is a trademark of ImmunoBioScience corp.

For research use only; not for use in diagnostic procedures. FOR IN VITRO LABORATORY USE ONLY

“In vitro laboratory products for research”

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC OR THERAPEUTIC PROCEDURES.